

## **POST-YIELD (Large strain)**

## Post-yield Strain Gauges YEF series ( & MGOBLET

These gauges are applicable to the measurement of large strain up to 10~15%. Also these withstand the repeated strain in elastic range (at strain level ±1500×10<sup>-6</sup> strain) like ordinary strain gauges. However, these are not applicable to the measurement of repeated strain in a large range. Integral leadwires using lead-free solder are available with option -F. This strain gauge is CE marked (compliant to RoHS2 Directive) and has joined to our "GOBLET" series.






	Gauge patter	rn		Туре		size(mm) Width	Backing : Length	size(mm) Width	Resist- ance Ω
Backing (Sauge width wid	cking length uge gth	=							
•Single axis			Single axis	YEFLAB-2	2	1.8	7	4	120
	YEFLAB-2	YEFLAB-5	Siligle axis	YEFLAB-5	5	2	10.5	4	120
●0° /90° 2-axis Plane type	· • • • • • • • • • • • • • • • • • • •		0°/90° 2-axis	YEFCAB-2	2	1.8	9.5	9.5	120
	YEFCAB-2	YEFCAB-5	Plane type	YEFCAB-5	5	2	13.5	13.5	120
●0° /45° /90° 3-axi Plane type	is 1 1 2		0°/45°/90° 3-axis	YEFRAB-2	2	1.8	9.5	9.5	120
	YEFRAB-2 YEFRAB-5		Plane type	YEFRAB-5	5	2	13.5	13.5	120

Minimum order quantity is 10 strain gauges. These strain gauges are available with integral leadwires attached. (made to order)

# Post-yield Strain Gauges YF series $C \in \mathbb{R}$

These gauges are applicable to the measurement of large strain up to 15 to 20%. These are not applicable to the measurement of repeated strain in elastic range as well as in large range.





	Gauge pattern	Туре	Gauge s Length		Backing : Length		
●Single axis							
	YFLA-10	YFLA-2	2	1.8	7.5	4	120
YFLA-2		YFLA-5	5	1.9	12	4	120
		YFLA-10	10	2.6	16.6	4.9	120
YFLA-5	YFLA-20	YFLA-20	20	1.8	26	3.7	120

Minimum order quantity is 10 strain gauges.


These strain gauges are available with integral leadwires attached. (made to order)




POST-YIELD (Large strain)

## 

These gauges are developed for the measurement of very large strain up to 30~40%. These are not applicable to the measurement of repeated strain in elastic range as well as in large range.





Gauge pattern				Туре	Gauge size(mm) Length Width		Backing size(mm) Length Width		Resist- ance Ω
●Single axis	s								
	YHFLA-2	1	<b>Q</b> (×3)	YHFLA-2	2	1.5	8	2.7	120
	YHFLA-5		Q (×3)	YHFLA-5	5	1.7	11	3	120
Minimum order quan These strain gauges		gauges. ith integral leadwires attached. (made to c	•						



### **Post-yield Strain Gauges**

### Dedicated leadwire recommended for YEF/YF/YHF series strain gauges

We supply various leadwires dedicated to strain gauges so as to meet our customers' requirements. Please refer to page 32 to 40 for the details of combination of a strain gauge and a leadwire

#### Type and designation of leadwires

Usage	Leadwire name	Operating temperature range of leadwire (°C)	Type number example		
General purpose (without temperature change)	Paralleled vinyl LJC-F	−20 ~+80	YEFLAB-2-3LJC-F YFLA-2-3LJC-F YHFLA-2-3LJC-F		
General use	3-wire paralleled vinyl LJCT-F	<b>−20 ~+80</b>	YEFLAB-2-3LJCT-F YFLA-2-3LJCT-F YHFLA-2-3LJCT-F		

### Important point

### ●Performance of YEF/YF/YHF

Series	Strain meas- urement	Fatigue limit at room temperature*1	Change of apparent strain due to cyclic loading of large strains*2	Self-temperature compensation	Applications
YEF	10~15%	5 x 10 <sup>5</sup> cycles	2000 x 10 <sup>-6</sup> strain/10 cycles	No	Measurement of 10 to 15% elongation     Measurement of repeated strain in elastic range.
YF	15~20%	1 x 10 <sup>2</sup> cycles	2000 x 10 <sup>-6</sup> strain/10 cycles	No	Measurement of 15 to 20% elongation
YHF	30~40%	2 x 10 <sup>4</sup> cycles	N/A	No	Measurement of 30 to 40% elongation
F	5%	1 x 10 <sup>6</sup> cycles	400 x 10 <sup>-6</sup> strain/10 cycles	Yes	Measurement of repeated strain in elastic range.     Measurement of 5% elongation


<sup>\*1:</sup> The number of repetitions at which the indicated strain value changes by 100x10-6 strain or more by applying repeated strain of approx. ±1,500x10-6 strain at 15Hz

### Adhesive for YEF/YF/YHF series gauges

These strain gauges should be bonded with CN or CN-Y adhesive. If measurement is made a few days or longer after the strain gauge bonding, the CN-Y should be used. Measurement of large strain is possible even after one year of bonding the strain gauge with the CN-Y adhesive, provided that the specimens are stored at room temperature without any unfavorable conditions (moisture, direct sunlight, etc.).

### ON adhesive variation with time

Though CN adhesive is normally used for large elongation strain measurement, the strain limit gradually decreases with the number of days following strain gauge installation. This variation with time occurs as a consequence of exposure to direct sunlight (UV), temperature and humidity, as well as the number of days after installation. The following shows an example of the results of testing performed by TML for the effects of adhesive variation with time. While these results show marked differences due to the exposure conditions of the test specimens (temperature and humidity), they also show that the strain limits for strain gauges decrease as time passes after installation. While this does not pose a problem in ordinary strain measurement, TML recommends that the measurement ends in 1 or 2 days after installation in the case of large elongation strain measurement. If the strain gauge is to be left for a long period after being installed, use the CN-Y adhesive.



#### Countermeasure in case there is a span between gauge installation and start of measurement

Store the test specimen with the attached strain gauge in a cool, dark and dry location. Use the CN-Y adhesive. (Refer to the instructions provided).

#### Repeatability of Post-Yield strain gauges

Post-Yield strain gauges can be used once to measure large elongation strain, but cannot be used for measurement of repeated large elongation strain. When repeated testing is performed in a strain range exceeding 5000x 10-6, the strain gauge experiences zero drift. Note that the amount of drift varies depending on factors such as the type of strain gauges and the level and frequncy of strain.

<sup>\*2:</sup> Change of indicated strain by applying a repeated strain of approx. ±10,000x10<sup>-6</sup> strain at a speed of 4 minutes per cycle.